

Prepare P750V1f calibration file.

New installation

First time the P750V1f is running the program will looking for a valid key.

When found you can see the serial number on left top of screen.

The serial number will be stored locally.

On all P750 systems there are a file where calibration values are stored if used. This file name is P750_XXXX.cfg, where the "XXXX" is the last 4 numbers of the serial number. If you use the same key as before you will not get a message that the program not can find this file.

Please copy the old P750_XXXX.cfg file from you system to the location where the new P750 program is stored or do following:

Start the P750 program. From start menu do "Ctl C" to activate the "Hardware Soft Calibration Menu".

Click Ok. This will generate a P750_XXXX.cfg file with default values for your serial number and key.

Please see user manual Appendix D for impedance calibration.